




Enable Log Frequency Scale on the DSA815

Date: 05/22/2015

Solution: Starting at firmware revision 00.01.12, the DSA815 spectrum analyzer can display the frequency scale in Log mode.

Press SPAN and set X Scale to Log





Headquarters

RIGOL TECHNOLOGIES, INC.
No.156,Cai He Village,
Sha He Town,
Chang Ping District, Beijing,
102206 P.R.China
Tel:+86-10-80706688
Fax:+86-10-80705070
Email: support@rigol.com

USA

RIGOL TECHNOLOGIES,USA INC.
102000 SW Allen Blvd, Suite C
Beaverton
OR 97005, USA
Toll free: 877-4-RIGOL-1
Office: (440) 232-4488
Fax: (216)-754-8107
Email: info@rigol.com

EUROPE

RIGOL TECHNOLOGIES GmbH
Lindbergh str. 4
82178 Puchheim
Germany
Tel: 0049- 89/89418950
Email: info-europe@rigoltech.com