

Product Overview

Ultra Spectrum is a PC application software developed by **RIGOL** for the DSA1000A, DSA1000 and DSA800 series spectrum analyzer. This software is compatible with the Windows XP, Windows Vista and Windows 7 operation systems.

Ultra Spectrum provides basic mode and advanced mode. Wherein, the basic mode provides the basic control and parameter configuration of the spectrum analyzer as well as the operation and processing of the data acquired from the spectrum analyzer; the advanced mode provides the basic mode functions together with more spectrum and data processing functions.

Product Orientation:

❖ Remote Control

This software is designed on the basis of the common standard drive VISA. You can build the communication between the software and spectrum analyzer via the USB or LAN interface to control the instrument remotely.

❖ Data Storage and Processing


This software provides powerful data processing functions (independent from the instrument). You can perform a series of operations (independent from the spectrum analyzer) on the spectral data acquired from the instrument to fulfill various signal measurement and study requirements.

❖ Instrument Function Extension and Enhancement

The special functions (such as the 3-D intensity graph) of the software extend and enhance the functions of the spectrum analyzer.



Product Features

- ❖ Mouse Operation, Easy-to-Use
- ❖ Various Marker Setting Functions
- ❖ Powerful Trace Display and Operation Functions
- ❖ Various Data Storage and Download Functions
- ❖ Peak and Valley Detect
- ❖ View History Spectrum Data
- ❖ Amplitude Correction Data Editing
- ❖ Various Advanced Measurement Functions
- ❖ SCPI Command List, Command Log



1. Mouse Operation, Easy-to-Use

You can use the mouse to modify the span, the amplitude range and the spectrum position, select the spectrum as well as zoom out/in the spectrum. The operation is direct and easy.



2. Various Marker Setting Functions

➤ One-click creation; up to 20 markers can be created.

➤ Provide various marker settings.



➤ Allow to set the marker value to the center frequency, start frequency, stop frequency, reference level and operand (for marker operation).

➤ The marker table displays the marker values.

Base Mark | Marks Operation

Create | Delete

Marker Name	X Value	Y Value
UserMarker	947500000.000	-50.964
UserMarker	952500000.000	-53.393
UserMarker	950000000.000	-55.060

➤ Support the addition, subtraction, multiplication and division operations between markers.

Base Mark | Marks Operation

NoMark | Do Operation

Addition

Marker Name	ResultX	ResultY
UserMarker	947500000.000	-53.744
UserMarker	952500000.000	-52.905
UserMarker	950000000.000	-53.433

3. Powerful Trace Display and Operation Functions

- Three kinds of trace display modes.

<input checked="" type="checkbox"/> Display Mode - Line (no Bar)
Display Mode - Vertical Bar
Display Mode - Horizontal Bar

- 7 basic operations and 12 user-defined operations; support easy and quick operation setting.

Peak
Valley
Max Hold
Min Hold
Freeze
Video Avg
Power Avg
Trace Setting

Trace Setting

Trace Name	Color	Operand
<input checked="" type="checkbox"/> Peak	Color	Original
<input checked="" type="checkbox"/> Original	Color	Original
<input type="checkbox"/> Valley	Color	Original
<input type="checkbox"/> MaxHold	Color	Original
<input type="checkbox"/> MinHold	Color	Original
<input type="checkbox"/> Freeze	Color	Original
<input type="checkbox"/> VideoAvg	Color	Original
<input type="checkbox"/> PowerAvg	Color	Original
<input type="checkbox"/> UTrace1	Color	Original
<input type="checkbox"/> UTrace2	Color	Original

Trace Setting


Trace Name	Color	Operand
<input type="checkbox"/> UTrace3	Color	Original
<input type="checkbox"/> UTrace4	Color	Original
<input type="checkbox"/> UTrace5	Color	Original
<input type="checkbox"/> UTrace6	Color	Original
<input type="checkbox"/> UTrace7	Color	Original
<input type="checkbox"/> UTrace8	Color	Original
<input type="checkbox"/> UTrace9	Color	Original
<input type="checkbox"/> UTrace10	Color	Original
<input type="checkbox"/> UTrace11	Color	Original
<input type="checkbox"/> UTrace12	Color	Original

Basic Operations

User-defined Operations

4. Various Data Storage and Download Functions

- Allow to store the instrument data in various formats.

Text (*.csv)	Picture (*.jpg, *.png or *.bmp)																					
<p>Trace(s) Data</p> <p>P/F Upper Limit-Original</p> <p>P/F Upper Limit-MaxHold</p> <p>P/F Upper Limit-MinHold</p> <p>P/F Upper Limit-Freeze</p> <p>P/F Lower Limit-Original</p> <p>P/F Lower Limit-MaxHold</p> <p>P/F Lower Limit-MinHold</p> <p>P/F Lower Limit-Freeze</p>	<p>Image with Information</p> <p>Save as Image</p> <p>Copy Image to Clipboard(Inverse)</p> <p>Copy Image to Clipboard(Normal)</p> <p>Inverse</p> <p>Normal</p>																					
<table border="1"> <thead> <tr> <th></th> <th>A</th> <th>B</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>Original-Frequency</td> <td>Original-Amplitude</td> </tr> <tr> <td>2</td> <td>0.00E+00</td> <td>-3.50E+00</td> </tr> <tr> <td>3</td> <td>8.42E+04</td> <td>-3.51E+00</td> </tr> <tr> <td>4</td> <td>1.68E+05</td> <td>-3.54E+00</td> </tr> <tr> <td>5</td> <td>2.53E+05</td> <td>-3.66E+00</td> </tr> <tr> <td>6</td> <td>3.37E+05</td> <td>-4.00E+00</td> </tr> </tbody> </table>		A	B	1	Original-Frequency	Original-Amplitude	2	0.00E+00	-3.50E+00	3	8.42E+04	-3.51E+00	4	1.68E+05	-3.54E+00	5	2.53E+05	-3.66E+00	6	3.37E+05	-4.00E+00	
	A	B																				
1	Original-Frequency	Original-Amplitude																				
2	0.00E+00	-3.50E+00																				
3	8.42E+04	-3.51E+00																				
4	1.68E+05	-3.54E+00																				
5	2.53E+05	-3.66E+00																				
6	3.37E+05	-4.00E+00																				

- State storage, loading and synchronization: allow to store the software state in *.txt format and recall the stored state when required; allow to synchronize the software state to the instrument.

Synchronize State to Instrument(Current State)
Synchronize State to Instrument(State File)
Record Current State(State File)

5. Peak and Valley Detect

- Drag the mouse to set the peak and valley limits and the software will search for the peaks and valleys that meet the conditions automatically.
- The peaks and valleys found are marked out on the trace with red dots and blue dots respectively and the coordinate values of the peaks and valleys found are displayed in table form.




- Allow to set the peak or valley value to the center frequency, start frequency, stop frequency or reference level.


Set to CenterFrequency
 Set to StartFrequency
 Set to StopFrequency
 Set to RefLevel

6. View History Spectrum Data

- Operate the localizer using the mouse to easily view the history spectral data and the spectrum window displays the trace at the localizer.



- Allow users to define the color of the waterfall graph and 3-D intensity graph.




- Waterfall graph:

- Display the 3-D ("frequency-time-amplitude") information of the signal.
- Display the marker point, maximum amplitude point and minimum amplitude point.
- Easy to view the amplitude distribution of the signal at different frequency to detect abnormal signals.
- Allow to save the waterfall graph as image (*.jpg, *.png or *.bmp) or text (*.dat or *.csv).
- Allow to load the waterfall data stored (*.dat).


Save as Image	
Save as Data	Raw Data (*.dat)
Copy Image to Clipboard	Conventional Data (*.csv)

➤ 3-D Intensity Graph:


- 3-D intensity display mode of the waterfall graph; provide 4 kinds of display modes including A-F-T, F-T, F-A and T-A.




A-F-T Mode



F-T Mode



F-A Mode




T-A Mode

- Allow to save the 3-D intensity graph as image (*.jpg, *.png or *.bmp).
- Flexible configuration of the display type of the 3-D intensity graph.


Visible All
✓ Visible Localizer
✓ Visible FT Grid
✓ Visible FA Grid
✓ Visible TA Grid
White Background
Display 3D Waterfall Only

7. Amplitude Correction Data Editing

- Various amplitude correction data editing methods:
- Edit the correction data in the data table (can be saved as *.csv file);
 - Load the *.csv file stored;
 - Load the edited amplitude correction data in the spectrum analyzer.




- Data preview function.



8. Various Advanced Measurement Functions


- Time Power: you can use the mouse to set the start line and stop line; the software draws the measurement values curve for you to easily view the variation tendency of the time power.




- Channel Power and Adjacent Channel Power:
 - Provide various main channel and adjacent channel creation methods including auto creation, creation on the basis of protocol (29 kinds of protocols) and manual editing method.
 - Display the measurement results in table and graph forms.




- Occupied Bandwidth and Emission Bandwidth: the software draws the measurement values curve for you to easily view the variation tendency of the parameters under test.




- C/N Ratio: you can use the mouse to set the noise bandwidth and the carrier bandwidth; the software draws the measurement values curve for you to easily view the variation tendency of the parameters under test.




- Harmonic Distortion: up to 10th order of harmonics can be measured; the measurement results are displayed in table and graph forms.




- Third Order Intermodulation Distortion



- Pass/Fail Test



- Provide various limits editing methods including manual editing, loading the stored limits file, importing spectral data as limits and generating limits by clicking the spectral window.
- Various test results display and storage modes.



- Create fail report in *.html format; the browser opens the fail report created automatically.

Fail Report --RIGOL® Spectrum Analyze
4:31 PM Thursday, June 14, 2012
Rigol Technologies, Inc.


Fail Number	Fail Frequency	Fail Amplitude	Fail Trace
100-1-1	15.600194E+6Hz	-83.3533dBm	FailTrace1
100-2-2	15.766916E+6Hz	-83.4011dBm	FailTrace1
100-3-3	16.050344E+6Hz	-83.5066dBm	FailTrace1
100-4-4	17.017333E+6Hz	-83.0195dBm	FailTrace1
100-5-5	17.450811E+6Hz	-82.3154dBm	FailTrace1
100-6-6	17.600861E+6Hz	-82.3679dBm	FailTrace1
100-7-7	17.700894E+6Hz	-84.4008dBm	FailTrace1
100-8-8	17.984322E+6Hz	-81.1916dBm	FailTrace1

Fail Number	Fail Frequency	Fail Amplitude	Limit Value	Delta
12-3-1	15.600194E+6Hz	-83.3533dBm	-83.0572dBm	0.2961dB
12-3-2	15.766916E+6Hz	-83.4011dBm	-82.9143dBm	0.4867dB
12-3-3	16.050344E+6Hz	-83.5066dBm	-82.6715dBm	0.8352dB
12-4-4	17.017333E+6Hz	-83.0195dBm	-81.8429dBm	1.1766dB
12-5-5	17.450811E+6Hz	-82.3154dBm	-81.4715dBm	0.8439dB
12-6-6	17.600861E+6Hz	-82.3679dBm	-81.3429dBm	1.0250dB
12-7-7	17.700894E+6Hz	-84.4008dBm	-81.2572dBm	3.1436dB

Fail Number	Fail Frequency	Fail Amplitude	Limit Value	Delta
11-3-1	16.000327E+6Hz	-84.3015dBm	-82.7143dBm	1.5872dB
11-3-2	16.733905E+6Hz	-86.4588dBm	-82.0857dBm	4.3730dB
11-3-3	16.900627E+6Hz	-86.0409dBm	-81.9429dBm	4.0980dB
11-4-4	16.983988E+6Hz	-84.1064dBm	-81.8715dBm	2.2349dB
11-5-5	17.317433E+6Hz	-83.5436dBm	-81.5857dBm	1.9579dB
11-6-6	17.400794E+6Hz	-82.6185dBm	-81.5143dBm	1.1042dB
11-7-7	17.484155E+6Hz	-81.4939dBm	-81.4429dBm	0.0510dB

9. SCPI Command List, Command Log

- Display the SCPI command list (including the time, command and sending state) of history operations.



- Create command log via the command list.



- Allow to copy or resend the commands in the list.

For detailed information of Ultra Spectrum, refer to the Ultra Spectrum Help Document.